

VETERINARY CHECKLIST and WORKSHEET

(A FACILITY checklist is required with this form)

Tube control number _____ Room number _____ ☐ Unit is in storage and not being used.

Control panel manufacturer and serial no. _____

Tubehead manufacturer and serial no. _____

Checklist

Choose Y for yes (compliant), N for no (noncompliant), N/A for not applicable and DNT for did not test.
(All reg references preceded by 0400-20-)

Y N N/A DNT

1. ☐ ☐ ☐ ☐ Switch at control panel and any remote switch labeled "CAUTION-RADIATION, THIS EQUIPMENT PRODUCES RADIATION WHEN ENERGIZED". (05-.111(11))
2. ☐ ☐ ☐ ☐ Proper technique chart posted and used. (04-.11(1)(e), 6-.04(4))
3. ☐ ☐ ☐ ☐ Operator remains in protected area during exposures or 2 meters away from primary beam. (06-.06(1)(a)7.)
4. ☐ ☐ ☐ ☐ Beam axis indicator for fixed units. (06-.05(2)(c)7., 06-.06(1)(a)1.)
5. ☐ ☐ ☐ ☐ Animal Holding Log maintained (holder's name, date, procedure). (06-.06(1)(b)1.(vii))
6. ☐ ☐ ☐ ☐ Distance from the x-ray field center to the light field center less than 2% of test SID. (06-.06(1)(a)1.)
7. ☐ ☐ ☐ ☐ Difference in dimensions of the light field and the x-ray field less than 2% of the test SID. (06-.06(1)(a)1.)
8. ☐ ☐ ☐ ☐ mm of Al HVL/Filtration adequate. (06-.06(1)(a)(2.)

Worksheet

(This section not required for RI use)

Settings:

kVp: _____

mA: _____

sec: _____

OR

mas: _____

Measurements:

kVp Measured: _____

HVL Measured: _____ **OR**

Total Filtration Measured: _____

[kVp<50, 0.5 mmAl: kVp 51-70, 1.5: kVp>71, 2.5]

Exposure Readings:

mR	Time

Source to Image Distance (SID): _____

2% of SID: _____

Notes