

Settleable Solids, SM 2540 F, 22nd edition (2011)

Initial Demonstration of Capability (DOC)

- Documentation (signed form) that analyst has read and understands all appropriate SOPs and Methods.

Method Detection Limit (MDL)

- NONE

Initial Calibration Verification (ICV)

- NONE

Method Blank

- NONE

Laboratory Fortified Blank (LFB)

- NONE

Duplicate –

- 2540 A.2. “To aid in quality assurance, analyze samples in duplicate.
- **Real people language – Analyze 2 samples for Sett. Solids.**
 - **For example, pour up 1000 mL of effluent into Imhoff then pour up another 1000 mL of effluent in another Imhoff. Wait 45 min, stir, wait 15 min, read. Figure RPD for both samples.**
 - **Target value is to get close to the first value and have a small RPD, less than 20%**
 - **Run on a 5% basis (1 for every 20 samples or once per month, whichever is more frequent), see batch size for more information**

Laboratory Fortified Matrix (LFM)/Laboratory Fortified Matrix Duplicate (LFMD)

- NONE

Control Charts

- NONE

Corrective Action - 1020 B.5., B.8., & B.15.

Batch Size

- For samples that need to be analyzed on a 5% basis (1 for every 20 samples or once per month, whichever is more frequent) follow these criteria:
 - If a permit stated that 3 analyses per week, we would allow for a duplicate to be analyzed at least once per month.
 - Pick a date and be consistent, the 1st of every month or the 1st Thursday of every month. Mark your calendar!!
 - If a permit stated 5 analyses per week, we would suggest twice a month.
 - Pick a date and be consistent, the 1st and 15th of every month or the 1st and 3rd Thursday of every month. Mark your calendar!!

Calculations

- RPD – relative percent differences for duplicates
 - = $\frac{\text{Difference between sample and duplicate}}{\text{Average of the sample and duplicate}} \times 100\%$